



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

**Garber Metrology Weighing Solutions &
Precision Calibration**

**520 E. Oregon Road
Lititz, PA 17543**

Fulfills the requirements of

ISO/IEC 17025:2017

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'J. Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 26 January 2027
Certificate Number: AC-1255



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

Garber Metrology Weighing Solutions & Precision Calibration

520 E. Oregon Road

Lititz, PA 17543

Zachary Saleet

717 393 1708 ext. 439

zsaleet@garbermetrology.com

www.garbermetrology.com

CALIBRATION

Valid to: **January 26, 2027**

Certificate Number: **AC-1255**

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---|---|---|
| DC Voltage – Source ¹ | Up to 220 mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V | 7.5 $\mu\text{V/V} + 0.4 \mu\text{V}$ 5 $\mu\text{V/V} + 0.7 \mu\text{V}$ 3.5 $\mu\text{V/V} + 2.5 \mu\text{V}$ 3.5 $\mu\text{V/V} + 4 \mu\text{V}$ 5 $\mu\text{V/V} + 40 \mu\text{V}$ 6.5 $\mu\text{V/V} + 0.4 \text{mV}$ | Comparison to Fluke 5730A Multiproduct Calibrator |
| DC Voltage – Measure ¹ | (0 to 100) mV 100 mV to 1 V (1 to 10) V (10 to 100) V (100 to 1 000) V | 7.5 $\mu\text{V/V} + 0.2 \mu\text{V}$ 2.9 $\mu\text{V/V} + 0.3 \mu\text{V}$ 2.9 $\mu\text{V/V} + 0.5 \mu\text{V}$ 4.3 $\mu\text{V/V} + 30 \text{nV}$ 4.4 $\mu\text{V/V} + 0.5 \text{mV}$ | Comparison to Fluke 8588A 8.5 Digit Multimeter |
| DC High Voltage – Measure ¹ | (1 to 10) kV (10 to 40) kV | 0.04 % of reading 0.39 % of reading + 10 V | Comparison to Vitretek 4700 High Voltage Meter, Fluke 5322A Safety Tester w/ High Voltage Probe |
| DC Current – Source ¹ | Up to 220 μA 220 μA to 2.2 mA (2.2 to 22) mA (22 to 220) mA 220 mA to 2.2 A | 40 $\mu\text{A/A} + 6 \text{nA}$ 35 $\mu\text{A/A} + 7 \text{nA}$ 35 $\mu\text{A/A} + 40 \text{nA}$ 45 $\mu\text{A/A} + 0.7 \mu\text{A}$ 80 $\mu\text{A/A} + 12 \mu\text{A}$ | Comparison to Fluke 5730A Multiproduct Calibrator |
| DC Current – Source ¹ | (2.2 to 2.999 9) A (0 to 10.999 9) A (11 to 20.5) A | 0.3 mA/A + 40 μA 0.4 mA/A + 0.5 mA 0.8 mA/A + 0.75 mA | Comparison to Fluke 5522A Multiproduct Calibrator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|--|--|
| DC Current – Measure ¹ | Up to 100 nA (0.1 to 1) μ A | 0.5 mA/A 69 μ A/A | Comparison to HP 3458A 8.5 Digit Multimeter |
| DC Current – Measure ¹ | (1 to 10) μ A (10 to 100) μ A 100 μ A to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1 A (1 to 10) A (10 to 30) A | 27 μ A/A + 0.4 nA 9.8 μ A/A + 0.4 nA 9.2 μ A/A + 4 nA 14 μ A/A + 40 nA 57 μ A/A + 0.1 μ A 0.13 mA/A + 1 μ A 0.23 mA/A + 4 μ A 0.55 mA/A + 0.44 mA | Comparison to Fluke 8588A 8.5 Digit Multimeter |
| DC Current – Source Clamp-on Meters ¹ | (10 to 50) A (50 to 100) A (100 to 500) A | 0.17 A 0.2 A 1.2 A | Comparison to Transmille EA002 Coil Adapter, Fluke 5522A Multiproduct Calibrator |
| Resistance – Source ¹ (Fixed Artifacts) | 0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω | 40 $\mu\Omega$ 95 $\mu\Omega$ 0.18 m Ω 0.23 m Ω 0.44 m Ω 1 m Ω 1.9 m Ω 6.5 m Ω 12.4 m Ω 65 m Ω 0.12 Ω 0.85 Ω 1.6 Ω 13 Ω 34 Ω 0.4 k Ω 0.9 k Ω 10 k Ω | Comparison to Fluke 5730A Multiproduct Calibrator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|--|---|
| Resistance – Measure ¹ | Up to 1 Ω (1 to 10) Ω (10 to 100) Ω 100 Ω to 1 kΩ (1 to 10) kΩ (10 to 100) kΩ 100 kΩ to 1 MΩ (1 to 10) MΩ (10 to 100) MΩ 100 MΩ to 1 GΩ | 17 μΩ/Ω + 4 μΩ 10.1 μΩ/Ω + 14 μΩ 9.2 μΩ/Ω + 50 μΩ 9.1 μΩ/Ω + 0.5 mΩ 9.2 μΩ/Ω + 5 mΩ 9.3 μΩ/Ω + 50 mΩ 10.6 μΩ/Ω + 1 Ω 19 μΩ/Ω + 0.1 kΩ 0.12 mΩ/Ω + 10 kΩ 1.3 mΩ/Ω + 1 MΩ | Comparison to Fluke 8588A 8.5 Digit Multimeter |
| Capacitance – Source ¹ (Simulation) | (220 to 399.9) pF (0.4 to 1.099 9) nF (1.1 to 3.299 9) nF (3.3 to 10.999 9) nF (11 to 32.999 9) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μF (1.1 to 3.299 99) μF (3.3 to 10.999 9) μF (11 to 32.999 9) μF (33 to 109.999) μF (110 to 329.999) μF (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.999 9) mF (11 to 32.999 9) mF (33 to 110) mF | 0.004 % of reading + 10 pF 0.004 % of reading + 10 pF 0.004 % of reading + 10 pF 0.002 % of reading + 10 pF 0.002 % of reading + 10 pF 0.002 % of reading + 10 pF 0.002 % of reading + 0.3 nF 0.002 % of reading + 1 nF 0.002 % of reading + 3 nF 0.002 % of reading + 10 nF 0.003 % of reading + 30 nF 0.004 % of reading + 0.1 μF 0.004 % of reading + 0.3 μF 0.004 % of reading + 1 μF 0.004 % of reading + 3 μF 0.004 % of reading + 10 μF 0.006 % of reading + 30 μF 0.009 % of reading + 0.1 mF | Comparison to Fluke 5522A Multiproduct Calibrator |
| Capacitance – Measure ¹ | Up to 1 nF (1 to 10) nF (10 to 100) nF 100 nF to 1 μF (1 to 10) μF (10 to 100) μF 100 μF to 1 mF (1 to 10) mF (10 to 100) mF | 1.8 nF/F + 1 pF 0.81 mF/F + 2 pF 0.49 mF/F + 10 pF 0.41 mF/F + 0.1 nF 0.42 mF/F + 1 nF 0.61 mF/F + 10 nF 0.62 mF/F + 0.1 μF 0.71 mF/F + 1 μF 0.71 mF/F + 10 μF | Comparison to Fluke 8588A 8.5 Digit Multimeter |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|----------------------------------|-------------------|---|---|
| AC Voltage – Source ¹ | Up to 2.2 mV | | Comparison to Fluke 5730A Multiproduct Calibrator |
| | (10 to 20) Hz | 0.24 mV/V + 4 μV | |
| | (20 to 40) Hz | 90 μV/V + 4 μV | |
| | 40 Hz to 20 kHz | 80 μV/V + 4 μV | |
| | (20 to 50) kHz | 0.2 mV/V + 4 μV | |
| | (50 to 100) kHz | 0.5 mV/V + 5 μV | |
| | (100 to 300) kHz | 1.1 mV/V + 10 μV | |
| | (300 to 500) kHz | 1.4 mV/V + 20 μV | |
| | 500 kHz to 1 MHz | 2.7 mV/V + 20 μV | |
| | (2.2 to 22) mV | | |
| | (10 to 20) Hz | 0.24 mV/V + 4 μV | |
| | (20 to 40) Hz | 90 μV/V + 4 μV | |
| | 40 Hz to 20 kHz | 80 μV/V + 4 μV | |
| | (20 to 50) kHz | 0.2 mV/V + 4 μV | |
| | (50 to 100) kHz | 0.5 mV/V + 5 μV | |
| | (100 to 300) kHz | 1.1 mV/V + 10 μV | |
| | (300 to 500) kHz | 1.4 mV/V + 20 μV | |
| | 500 kHz to 1 MHz | 2.7 mV/V + 20 μV | |
| | (22 to 220) mV | | |
| | (10 to 20) Hz | 0.24 mV/V + 12 μV | |
| | (20 to 40) Hz | 90 μV/V + 7 μV | |
| | 40 Hz to 20 kHz | 60 μV/V + 7 μV | |
| | (20 to 50) kHz | 0.12 mV/V + 7 μV | |
| | (50 to 100) kHz | 0.31 mV/V + 17 μV | |
| (100 to 300) kHz | 0.66 mV/V + 20 μV | | |
| (300 to 500) kHz | 1.4 mV/V + 25 μV | | |
| 500 kHz to 1 MHz | 2.7 mV/V + 45 μV | | |
| 220 mV to 2.2 V | | | |
| (10 to 20) Hz | 0.24 mV/V + 40 μV | | |
| (20 to 40) Hz | 90 μV/V + 15 μV | | |
| 40 Hz to 20 kHz | 40 μV/V + 8 μV | | |
| (20 to 50) kHz | 70 μV/V + 10 μV | | |
| (50 to 100) kHz | 90 μV/V + 30 μV | | |
| (100 to 300) kHz | 0.34 mV/V + 80 μV | | |
| (300 to 500) kHz | 1 mV/V + 0.2 mV | | |
| 500 kHz to 1 MHz | 1.7 mV/V + 0.3 mV | | |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-----------------------------------|--------------------|---|---|
| AC Voltage – Source ¹ | (2.2 to 22) V | | Comparison to Fluke 5730A Multiproduct Calibrator |
| | (10 to 20) Hz | 0.24 mV/V + 0.4 mV | |
| | (20 to 40) Hz | 90 μV/V + 0.15 mV | |
| | 40 Hz to 20 kHz | 40 μV/V + 50 μV | |
| | (20 to 50) kHz | 70 μV/V + 0.1 mV | |
| | (50 to 100) kHz | 80 μV/V + 0.2 mV | |
| | (100 to 300) kHz | 0.25 mV/V + 0.6 mV | |
| | (300 to 500) kHz | 1 mV/V + 2 mV | |
| | 500 kHz to 1 MHz | 1.5 mV/V + 3.2 mV | |
| | (22 to 220) V | | |
| | (10 to 20) Hz | 0.24 mV/V + 4 mV | |
| | (20 to 40) Hz | 90 μV/V + 1.5 mV | |
| | 40 Hz to 20 kHz | 50 μV/V + 0.6 mV | |
| | (20 to 50) kHz | 80 μV/V + 1 mV | |
| | (50 to 100) kHz | 0.15 mV/V + 2.5 mV | |
| (100 to 300) kHz | 0.9 mV/V + 16 mV | | |
| (300 to 500) kHz | 4.4 mV/V + 40 mV | | |
| 500 kHz to 1 MHz | 8 mV/V + 80 mV | | |
| AC Voltage – Measure ¹ | (220 to 1 100) V | | Comparison to Fluke 8588A 8.5 Digit Multimeter |
| | (15 to 50) Hz | 0.3 mV/V + 16 mV | |
| | 50 Hz to 1 kHz | 70 μV/V + 3.5 mV | |
| | Up to 10 mV | | |
| | 1 Hz to 2 kHz | 0.29 mV/V + 1.1 μV | |
| | (2 to 10) kHz | 0.37 mV/V + 1.1 μV | |
| | (10 to 30) kHz | 0.38 mV/V + 1.1 μV | |
| | (30 to 100) kHz | 3 mV/V + 1.1 μV | |
| | (100 to 300) kHz | 10 mV/V + 4 μV | |
| | 300 kHz to 1 MHz | 20 mV/V + 4 μV | |
| | (10 to 100) mV | | |
| | 1 Hz to 2 kHz | 88 μV/V + 0.5 μV | |
| | (2 to 10) kHz | 0.13 mV/V + 0.5 μV | |
| | (10 to 30) kHz | 0.23 mV/V + 1.0 μV | |
| | (30 to 100) kHz | 0.53 mV/V + 5.0 μV | |
| (100 to 300) kHz | 2.1 mV/V + 30 μV | | |
| 300 kHz to 1 MHz | 11 mV/V + 0.1 mV | | |
| (1 to 2) MHz | 15.4 mV/V + 0.5 mV | | |
| (2 to 4) MHz | 41 mV/V + 1 mV | | |
| (4 to 8) MHz | 84 mV/V + 1 mV | | |
| (8 to 10) MHz | 0.16 V/V + 1 mV | | |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|------------------------------|---|--|
| AC Voltage – Measure ¹ | 100 mV to 1 V | | Comparison to Fluke 8588A 8.5 Digit Multimeter |
| | 1 Hz to 2 kHz | 76 μ V/V + 5 μ V | |
| | (2 to 10) kHz | 0.12 mV/V + 5 μ V | |
| | (10 to 30) kHz | 0.23 mV/V + 10 μ V | |
| | (30 to 100) kHz | 0.53 mV/V + 50 μ V | |
| | (100 to 300) kHz | 2.1 mV/V + 0.3 mV | |
| | 300 kHz to 1 MHz | 10 mV/V + 1 mV | |
| | (1 to 2) MHz | 15 mV/V + 5 mV | |
| | (2 to 4) MHz | 40 mV/V + 10 mV | |
| | (4 to 8) MHz | 82 mV/V + 10 mV | |
| | (8 to 10) MHz | 0.15 V/V + 10 mV | |
| | (1 to 10) V | | |
| | 1 Hz to 2 kHz | 76 μ V/V + 50 μ V | |
| | (2 to 10) kHz | 0.12 mV/V + 50 μ V | |
| | (10 to 30) kHz | 0.13 mV/V + 0.1 mV | |
| | (30 to 100) kHz | 0.53 mV/V + 0.5 mV | |
| | (100 to 300) kHz | 2.1 mV/V + 3 mV | |
| | 300 kHz to 1 MHz | 10 mV/V + 10 mV | |
| | (1 to 2) MHz | 15 mV/V + 50 mV | |
| | (2 to 4) MHz | 40 mV/V + 0.1 V | |
| (4 to 8) MHz | 82 mV/V + 0.1 V | | |
| (8 to 10) MHz | 0.15 V/V + 0.1 V | | |
| (10 to 100) V | | | |
| 1 Hz to 2 kHz | 90 μ V/V + 0.5 mV | | |
| (2 to 10) kHz | 0.11 mV/V + 0.5 mV | | |
| (10 to 30) kHz | 0.23 mV/V + 1 mV | | |
| (30 to 100) kHz | 0.59 mV/V + 5 mV | | |
| (100 to 300) kHz | 3.7 mV/V + 50 mV | | |
| 300 kHz to 1 MHz | 10.1 mV/V + 0.5 V | | |
| (100 to 1 000) V | | | |
| 1 Hz to 2 kHz | 0.11 mV/V + 25 mV | | |
| (2 to 10) kHz | 0.11 mV/V + 25 mV | | |
| (10 to 30) kHz | 0.23 mV/V + 25 mV | | |
| (30 to 100) kHz | 0.59 mV/V + 0.1 V | | |
| AC High Voltage – Measure ¹ | (50 to 60) Hz (1 to 5) kV | 0.15 % of reading | Comparison to Vitrek 4700 High Voltage Meter |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|----------------------------------|--|---|---|
| AC Current – Source ¹ | Up to 220 μ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz 220 μ A to 2.2 mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz 220 mA to 2.2 A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz | 0.25 mA/A + 16 nA 0.16 mA/A + 10 nA 0.1 mA/A + 8 nA 0.28 mA/A + 12 nA 1.1 mA/A + 65 nA 0.25 mA/A + 40 nA 0.16 mA/A + 35 nA 0.1 mA/A + 35 nA 0.2 mA/A + 0.11 μ A 1.1 mA/A + 0.65 μ A 0.25 mA/A + 0.4 μ A 0.16 mA/A + 0.35 μ A 0.1 mA/A + 0.35 μ A 0.2 mA/A + 0.55 μ A 1.1 mA/A + 5 μ A 0.25 mA/A + 4 μ A 0.16 mA/A + 3.5 μ A 0.1 mA/A + 2.5 μ A 0.2 mA/A + 3.5 μ A 1.1 mA/A + 10 μ A 0.24 mA/A + 35 μ A 0.45 μ A/A + 80 μ A 7 mA/A + 0.16 mA | Comparison to Fluke 5730A Multiproduct Calibrator |
| AC Current – Source ¹ | (1.1 to 2.999) A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (3 to 10.999) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz (11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz | 0.14 % of reading + 0.1 mA 0.04 % of reading + 0.1 mA 0.5 % of reading + 1 mA 1.9 % of reading + 5 mA 0.05 % of reading + 2 mA 0.08 % of reading + 2 mA 2.3 % of reading + 2 mA 0.09 % of reading + 5 mA 0.12 % of reading + 5 mA 2.3 % of reading + 5 mA | Comparison to Fluke 5522A Multiproduct Calibrator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|-----------------------|---|--|
| AC Current – Measure ¹ | Up to 10 μ A | | Comparison to Fluke 8588A 8.5 Digit Multimeter |
| | 1Hz to 2 kHz | 2 mA/A + 2.5 nA | |
| | (2 to 10) kHz | 2 mA/A + 2.5 nA | |
| | (10 to 30) kHz | 2 mA/A + 2.5 nA | |
| | (10 to 100) μ A | | |
| | 1 Hz to 2 kHz | 0.28 mA/A + 5 nA | |
| | (2 to 10) kHz | 0.53 mA/A + 5 nA | |
| | (10 to 30) kHz | 0.74 mA/A + 5 nA | |
| | (30 to 100) kHz | 4 mA/A + 10 nA | |
| | 100 μ A to 1 mA | | |
| | 1 Hz to 2 kHz | 0.28 mA/A + 50 nA | |
| | (2 to 10) kHz | 0.53 mA/A + 50 nA | |
| | (10 to 30) kHz | 0.74 mA/A + 50 nA | |
| | (30 to 100) kHz | 4 mA/A + 0.1 μ A | |
| | (1 to 10) mA | | |
| | 1 Hz to 2 kHz | 0.28 mA/A + 0.5 μ A | |
| | (2 to 10) kHz | 0.53 mA/A + 0.5 μ A | |
| | (10 to 30) kHz | 0.74 mA/A + 0.5 μ A | |
| | (30 to 100) kHz | 4 mA/A + 1 μ A | |
| | (10 to 100) mA | | |
| 1 Hz to 2 kHz | 0.28 mA/A + 5 μ A | | |
| (2 to 10) kHz | 0.52 mA/A + 5 μ A | | |
| (10 to 30) kHz | 0.74 mA/A + 5 μ A | | |
| 100 mA to 1 A | | | |
| 1 Hz to 2 kHz | 0.3 mA/A + 0.1 mA | | |
| (2 to 10) kHz | 0.55 mA/A + 0.1 mA | | |
| (10 to 30) kHz | 0.79 mA/A + 0.1A | | |
| (1 to 10) A | | | |
| 10 Hz to 2 kHz | 0.84 mA/A + 1 mA | | |
| (2 to 10) kHz | 0.84 mA/A + 1mA | | |
| (10 to 30) A | | | |
| 10 Hz to 2 kHz | 0.8 mA/A + 12 mA | | |
| 2 kHz to 10 kHz | 1.2 mA/A + 12 mA | | |
| AC Current – Source Clamp-on Meters ¹ | 60 Hz | | Comparison to Transmille EA002 Coil Adapter, Fluke 5522A Multiproduct Calibrator |
| | (10 to 100) A | 0.19 A | |
| | (100 to 500) A | 1.19 A | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|----------------------------|----------|---|---|
| Oscilloscopes ¹ | | | |
| Amplitude – DC | | | |
| 50 Ω load | 0 V | 12 mV | |
| | 6 V | 12 mV | |
| 1 MΩ load | 0 V | 12 mV | |
| | 66 V | 43 mV | |
| | 130 V | 76 mV | |
| Amplitude – Square wave | | | |
| into 50 Ω load | 0.1 Vp-p | | |
| | 10 kHz | 5.9 mV | |
| | 1 Vp-p | | |
| | 10 kHz | 5.9 mV | |
| | 5 Vp-p | | |
| | 10 kHz | 15 mV | |
| into 1 MΩ load | 0.1 Vp-p | | |
| | 10 kHz | 0.43 mV | |
| | 1 Vp-p | | |
| | 10 kHz | 2.6 mV | |
| | 10 Vp-p | | |
| | 10 kHz | 6.4 mV | |
| Leveled Sine Flatness | | | |
| (relative to 50 kHz) | | | |
| into 50 Ω load | 10 mVp-p | | |
| | 50 kHz | 0.58 mV | |
| | 30 mVp-p | | |
| | 100 kHz | 2.1 mV | |
| | 300 MHz | 2.6 mV | |
| | 600 MHz | 5 mV | |
| | 5 V p-p | | |
| | 50 kHz | 0.36 mV | |
| | 100 kHz | 1.7 mV | |
| | 300 MHz | 2.3 mV | |
| | 600 MHz | 4.6 mV | |
| | | | Comparison to Fluke 5522A Multiproduct Calibrator with 1 100 MHz Scope Option |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment | |
|---|-----------------------------------|---|---|-------|
| Oscilloscopes ¹ Rise Time | 1 MHz 1 Vp-p Up to 400 ps | 8.3 ps | Comparison to Fluke 5522A Multiproduct Calibrator with 1 100 MHz Scope Option | |
| | 10 MHz 0.5 Vp-p Up to 400 ps | 8.2 ps | | |
| | 1 Vp-p Up to 400 ps | 8.2 ps | | |
| | Time Markers | 2 ns | | 10 ps |
| | | 20 ms | | 12 μs |
| | 50 ms | 10 μs | | |
| | 5 s | 40 ms | | |
| Electrical Simulation of RTD Indicating Devices ¹ | Cu 427, 10 Ω (-100 to 260) °C | 0.23 °C | Comparison to Fluke 5522A Multiproduct Calibrator | |
| | Pt 385, 100 Ω (-200 to -80) °C | 0.04 °C | | |
| | (-80 to 0) °C | 0.04 °C | | |
| | (0 to 100) °C | 0.05 °C | | |
| | (100 to 300) °C | 0.07 °C | | |
| | (300 to 400) °C | 0.08 °C | | |
| | (400 to 630) °C | 0.09 °C | | |
| | (630 to 800) °C | 0.18 °C | | |
| | Pt 385, 200 Ω (-200 to -80) °C | 0.03 °C | | |
| | (-80 to 0) °C | 0.03 °C | | |
| | (0 to 100) °C | 0.03 °C | | |
| | (100 to 260) °C | 0.04 °C | | |
| | (260 to 300) °C | 0.09 °C | | |
| | (300 to 400) °C | 0.1 °C | | |
| | (400 to 600) °C | 0.11 °C | | |
| (600 to 630) °C | 0.12 °C | | | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|-------------------|---|---|
| Electrical Simulation of RTD Indicating Devices ¹ | Pt 385, 500 Ω | | Comparison to Fluke 5522A Multiproduct Calibrator |
| | (-200 to -80) °C | 0.03 °C | |
| | (-80 to 0) °C | 0.04 °C | |
| | (0 to 100) °C | 0.04 °C | |
| | (100 to 260) °C | 0.05 °C | |
| | (260 to 300) °C | 0.06 °C | |
| | (300 to 400) °C | 0.06 °C | |
| | (400 to 600) °C | 0.07 °C | |
| | (600 to 630) °C | 0.09 °C | |
| | Pt 385, 1 000 Ω | | |
| | (-200 to -80) °C | 0.02 °C | |
| | (-80 to 0) °C | 0.02 °C | |
| | (0 to 100) °C | 0.03 °C | |
| | (100 to 260) °C | 0.04 °C | |
| | (260 to 300) °C | 0.05 °C | |
| | (300 to 400) °C | 0.05 °C | |
| | (400 to 600) °C | 0.05 °C | |
| | (600 to 630) °C | 0.18 °C | |
| | Pt 3926, 100 Ω | | |
| | (-200 to -80) °C | 0.04 °C | |
| | (-80 to 0) °C | 0.04 °C | |
| | (0 to 100) °C | 0.05 °C | |
| | (100 to 300) °C | 0.07 °C | |
| | (300 to 400) °C | 0.08 °C | |
| | (400 to 630) °C | 0.09 °C | |
| | Pt 3916, 100 Ω | | |
| | (-200 to -190) °C | 0.19 °C | |
| | (-190 to -80) °C | 0.03 °C | |
| (-80 to 0) °C | 0.04 °C | | |
| (0 to 100) °C | 0.05 °C | | |
| (100 to 260) °C | 0.05 °C | | |
| (260 to 300) °C | 0.06 °C | | |
| (300 to 400) °C | 0.07 °C | | |
| (400 to 600) °C | 0.08 °C | | |
| (600 to 630) °C | 0.18 °C | | |
| PtNi 385, 120 Ω | | | |
| (-80 to 0) °C | 0.06 °C | | |
| (0 to 100) °C | 0.06 °C | | |
| (100 to 260) °C | 0.11 °C | | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---------------------|---|---|
| Electrical Simulation of Thermocouple – Measure/Source ¹ | Type B | | Comparison to Fluke 5522A Multiproduct Calibrator |
| | (600 to 800) °C | 0.34 °C | |
| | (800 to 1 000) °C | 0.26 °C | |
| | (1 000 to 1 550) °C | 0.23 °C | |
| | (1 550 to 1 820) °C | 0.26 °C | |
| | Type C | | |
| | (0 to 150) °C | 0.23 °C | |
| | (150 to 650) °C | 0.2 °C | |
| | (650 to 1 000) °C | 0.24 °C | |
| | (1 000 to 1 800) °C | 0.39 °C | |
| | (1 800 to 2 316) °C | 0.65 °C | |
| | Type E | | |
| | (-250 to -100) °C | 0.39 °C | |
| | (-100 to -25) °C | 0.12 °C | |
| | (-25 to 350) °C | 0.11 °C | |
| | (350 to 650) °C | 0.12 °C | |
| | (650 to 1 000) °C | 0.16 °C | |
| | Type J | | |
| | (-210 to -100) °C | 0.21 °C | |
| | (-100 to -30) °C | 0.12 °C | |
| | (-30 to 150) °C | 0.11 °C | |
| | (150 to 760) °C | 0.13 °C | |
| | (760 to 1 200) °C | 0.18 °C | |
| Type K | | | |
| (-200 to -100) °C | 0.26 °C | | |
| (-100 to -25) °C | 0.14 °C | | |
| (-25 to 120) °C | 0.12 °C | | |
| (120 to 1 000) °C | 0.2 °C | | |
| (1 000 to 1 372) °C | 0.31 °C | | |
| Type L | | | |
| (-200 to -100) °C | 0.29 °C | | |
| (-100 to 800) °C | 0.2 °C | | |
| (800 to 900) °C | 0.13 °C | | |
| Type N | | | |
| (-200 to -100) °C | 0.31 °C | | |
| (-100 to -25) °C | 0.17 °C | | |
| (-25 to 120) °C | 0.15 °C | | |
| (120 to 410) °C | 0.14 °C | | |
| (410 to 1 300) °C | 0.21 °C | | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---------------------|---|---|
| Electrical Simulation of Thermocouple – Measure/Source ¹ | Type R | | Comparison to Fluke 5522A Multiproduct Calibrator |
| | (0 to 250) °C | 0.44 °C | |
| | (250 to 400) °C | 0.27 °C | |
| | (400 to 1 000) °C | 0.26 °C | |
| | (1 000 to 1 767) °C | 0.31 °C | |
| | Type S | | |
| | (0 to 250) °C | 0.36 °C | |
| | (250 to 1 000) °C | 0.28 °C | |
| | (1 000 to 1 400) °C | 0.29 °C | |
| | (1 400 to 1 767) °C | 0.36 °C | |
| | Type T | | |
| | (-250 to -150) °C | 0.49 °C | |
| | (-150 to 0) °C | 0.19 °C | |
| (0 to 120) °C | 0.12 °C | | |
| (120 to 400) °C | 0.11 °C | | |
| Type U | | | |
| (-200 to 0) °C | 0.43 °C | | |
| (0 to 600) °C | 0.21 °C | | |

Length – Dimensional Metrology

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---------------------------------------|--|---|---|
| Gage Blocks ² | (0.01 to 4) in | $(3.1 + 1L) \mu\text{in}$ | Comparison to Pratt & Whitney LMU-2130 Comparator, Grade 1 Gage Blocks |
| Long Gage Blocks ² | (4 to 20) in | $(2.5 + 1.2L) \mu\text{in}$ | Comparison to Pratt & Whitney LMU-1000M Comparator, Grade 0 Gage Blocks |
| Thread Measuring Wires (4 to 120) TPI | (0.004 81 to 0.144 352) in | 8.8 μin | Comparison to Pratt & Whitney LMU-2130 Comparator, Grade 1 Gage Blocks |
| Plain Plugs/Pin Gages | (0.004 to 1) in (1 to 4) in (4 to 12) in | 6.8 μin 9.7 μin 19 μin | Comparison to Pratt & Whitney LMU-2130 Comparator, Grade 1 Gage Blocks |
| Z-Mike Laser Micrometer | Up to 1 in | 31 μin | Comparison to Class XXX Pins |

Length – Dimensional Metrology

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|---|--|
| Pins | Up to 1 in | 43 μ in | Comparison to Z-Mike Laser Micrometer |
| Thread Plugs (Pitch Diameter) | Up to 1 in (1 to 3) in (3 to 7.5) in | 110 μ in 112 μ in 123 μ in | Comparison to Pratt & Whitney LMU-2130 Comparator, Grade 1 Gage Blocks, Thread Measuring Wires |
| Thread Plugs (Pitch Diameter) | (7.5 to 12) in | 142 μ in | Comparison to Pratt & Whitney LMU-2130 Comparator, Grade 1 Gage Blocks, Thread Measuring Wires |
| NPT Thread Plugs (Pitch Diameter) | Up to 1 in (1 to 3) in (3 to 6) in | 110 μ in 112 μ in 123 μ in | Comparison to Pratt & Whitney LMU-2130 Comparator, Grade 1 Gage Blocks, Thread Measuring Wires |
| Adjustable Thread Rings | Up to 1 in (1 to 4) in (4 to 8) in | Uncertainty is calculated using the Master Thread Setting Plug Uncertainty. | Tactile fit to a Master Thread Setting Plug |
| Plain Rings | (0.04 to 1) in (1 to 4) in (4 to 8) in (8 to 12) in | 12 μ in 15 μ in 25 μ in 32 μ in | Comparison to Pratt & Whitney LMU-2130 Comparator, Class XXX Plain Rings |
| Plain Rings | (12 to 18) in | 33 μ in | Comparison to Pratt & Whitney LMU-1000M Comparator, Class XXX Plain Rings |
| Micrometers ^{1,2} (OD, ID, Bore, Depth) (0.001 in resolution) (0.000 1 in resolution) (0.000 05 in resolution) | (0.05 to 72) in (0.05 to 72) in (0.05 to 24) in | (580 + 1.1L) μ in (58 + 3.6L) μ in (29 + 3.4L) μ in | Comparison to Grade 2 Gage Blocks, Optical Flat |
| Calipers ^{1,2} (Dial, Vernier, & Digital) (0.001 in resolution) (0.000 5 in resolution) | (0.05 to 120) in (0.05 to 120) in | (580 + 1.7L) μ in (290 + 2.6L) μ in | Comparison to Grade 2 Gage Blocks, Optical Flat |
| Indicator Calibrators | Up to 1 in | 59 μ in | Comparison to Grade 2 Gage Blocks, Optical Flat |

Length – Dimensional Metrology

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---|--|--|
| Height Gages ¹ | Up to 12 in (12 to 48) in | 600 μin 615 μin | Comparison to Grade 2 Gage Blocks, Optical Flat |
| Dial and Digital Indicators ¹ (0.001 in resolution) (0.000 1 in resolution) (0.000 05 in resolution) (0.000 02 in resolution) (0.000 01 in resolution) | Up to 6 in Up to 0.5 in Up to 0.05 in Up to 0.02 in Up to 0.01 in | 290 μin 140 μin 58 μin 34 μin 14 μin | Comparison to Grade 2 Gage Blocks, Indicator Calibrator |
| Surface Plates ^{1,2} | | | In accordance with Fed Spec GGG-P-463 using Planekator, Straight Indicators |
| Overall Flatness | Up to 16.97 <i>DL</i> Up to 21.63 <i>DL</i> Up to 60 <i>DL</i> Up to 161 <i>DL</i> | 52 μin 56 μin 240 μin 240 μin | |
| Local Area Flatness (Repeat Reading) | Up to 0.000 02 in | 55 μin | Repeat-o-Meter |
| Length Standards ² | (0.5 to 4) in (5 to 20) in (21 to 42) in (43 to 96) in | (3.9 + 1.5 <i>L</i>) μin (29 + 0.9 <i>L</i>) μin (18 + 1 <i>L</i>) μin (9.7 + 2.1 <i>L</i>) μin | Comparison to Pratt & Whitney LMU-2130 Comparator, Pratt & Whitney LMU-1000 Comparator Grade 1 Gage Blocks, Electronic Indicator |
| Digital Levels ¹ | 0° 15° 30° 45° 90° | 0.06° 0.13° 0.15° 0.15° 0.13° | Comparison to Grade 2 Gage Blocks, Sine Bar |
| Parallels | Up to 4 in | 10 μin | Comparison to Pratt & Whitney LMU-2130 Comparator, Grade 1 Gage Blocks |
| Optical Comparators ^{1,2} X, Y Axis Length | Up to 12 in | (120 + 5.3 <i>L</i>) μin | Comparison to Glass Scale |
| Angle | Up to 360° | 0.06° | Angle Blocks |

Length – Dimensional Metrology

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|------------------------------------|---|--|
| Video Measuring Machines ^{1,2} X,Y Axis Length | Up to 12 in | $(68 + 2.1L) \mu\text{in}$ | Comparison to Glass Scale |
| Z Axis Length | Up to 4 in | $(210 + 17L) \mu\text{in}$ | Step Gage |
| Angle | Up to 360° | 0.02° | Angle Blocks |
| Roughness Specimen | Ra: (0.8 to 500) μin | 0.34 % of reading + 1.1 μin | Comparison to Mitutoyo Surface Roughness Tester, Contour Measuring Machine |

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|--|---|
| Low Pressure Devices ¹ (Pneumatic Gauge Pressure) | (0.01 to 30) inH ₂ O | 0.096 % of reading + 0.001 2 inH ₂ O | Comparison to Additel Pressure Module |
| Hydraulic Pressure Devices ¹ (Gauge Pressure) | (10 to 16 000) psig (16 000 to 20 000) psig | 0.009 4 % of reading 0.015 % of reading + 0.4 psi | Comparison to Hydraulic Deadweight Testers Fluke P3125 Fluke 3116 |
| Vacuum Devices ¹ (Pneumatic) | (-15 to -0.1) psig | 0.003 8 psi | Comparison to Additel 673 Digital Pressure Calibrator |
| Torque Tools ¹ | (20 to 200) ozf·in (5 to 50) lbf·in (40 to 400) lbf·in (100 to 1 000) lbf·in (25 to 250) lbf·ft (50 to 500) lbf·ft (100 to 1 000) lbf·ft (200 to 2 000) lbf·ft | 0.11 % of reading + 1.5 ozf·in 0.067 % of reading + 0.48 lbf·in 0.17 % of reading + 1 lbf·in 0.22 % of reading + 1.8 lbf·in 0.2 % of reading + 0.49 lbf·ft 0.27 % of reading + 2.6 lbf·ft 0.21 % of reading + 7.2 lbf·ft 0.056 % of reading + 37 lbf·ft | Comparison to CDI Sure-test 5000-ST Torque Calibration System |
| Torque Calibrator | (4 to 50) lbf·in (30 to 400) lbf·in (100 to 1 000) lbf·in (20 to 250) lbf·ft | 0.01 lbf·in 0.03 lbf·in 0.13 lbf·in 0.04 lbf·ft | Comparison to Torque Arms, Class F Weights |



ANSI National Accreditation Board

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|--|--|
| Durometers Spring Force Only Types A, B, E, O Types C, D, DO | (0 to 8.05) N (0 to 44.45) N | 0.089 N 0.53 N | Partial Direct Verification per internal procedure CP-0053 using Triple Beam Balance |
| Class F Mass Determination | (1 to 2) g (5 to 100) g 200 g 500 g 1 000 g (2 000 to 5 000) g (0.001 to 0.002) lb (0.005 to 0.2) lb (0.5 to 10) lb (10 to 50) lb | 0.3 mg 0.4 mg 13 mg 22 mg 33 mg 56 mg 0.000 000 5 lb 0.000 014 lb 0.000 3 lb 0.001 6 lb | Balance, ASTM E617 Class 3 Weights; NIST HB 105-1 |
| Balances and Scales ¹ 0.1 mg resolution | Up to 10 g Up to 200 g | 0.2 mg 0.3 mg | ASTM E617 Class 0 weights and NIST Handbook 44 utilized in the calibration of the weighing device. |
| Balances and Scales ¹ 0.1 mg resolution | (200 to 600) g (600 to 6 000) g | 15 mg 22 mg | ASTM E617 Class 0 weights and NIST Handbook 44 utilized in the calibration of the weighing device. |
| Balances and Scales ¹ 0.1 g resolution | Up to 1.2 kg (1.2 to 2) kg (2 to 6) kg (5 to 30) kg | 0.1 g 0.1 g 0.2 g 0.2 g | NIST Class F weights and NIST Handbook 44 utilized in the calibration of the weighing device. |
| Balances and Scales ¹ 0.000 2 lb resolution 0.000 5 lb resolution 0.001 lb resolution 0.005 lb resolution 0.002 lb resolution 0.005 lb resolution | Up to 2 lb Up to 5 lb Up to 10 lb Up to 20 lb Up to 25 lb Up to 50 lb | 0.000 4 lb 0.001 lb 0.002 lb 0.01 lb 0.004 lb 0.01 lb | NIST Class F weights and NIST Handbook 44 utilized in the calibration of the weighing device. |

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---|--|---|
| Balances and Scales ¹ 0.01 lb resolution 0.05 lb resolution 0.05 lb resolution 0.2 lb resolution 0.5 lb resolution 1 lb resolution 2 lb resolution 20 lb resolution | Up to 100 lb Up to 150 lb Up to 500 lb Up to 1 000 lb Up to 3 000 lb Up to 5 000 lb Up to 20 000 lb Up to 200 000 lb | 0.03 lb 0.1 lb 0.1 lb 0.3 lb 0.6 lb 1.3 lb 2.6 lb 27 lb | NIST Class F weights and NIST Handbook 44 utilized in the calibration of the weighing device. |
| Force Measuring Devices | Up to 1 000 lbf | 0.1 % of reading | Comparison to Morehouse Load Cells with Readout |

Thermodynamic

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---------------------------------|--|--|
| Temperature – Source ¹ (Temperature Measuring Devices) | (-5 to 140) °C | 0.24 °C | Comparison to Hart Scientific 9105 Drywell |
| Temperature – Source ¹ (Temperature Measuring Devices) | (0 to 260) °C | 0.69 °C | Comparison to Hart 6102 Micro-bath |
| Thermo-hygrometers Temperature Humidity | (0 to 180) °C (30 to 90) %RH | 0.5 % of reading + 0.15 °C 0.5 % of reading + 0.9 %RH | Comparison to Vaisala MI70/HMP75 Temp/Humidity Indicator with Probe |
| Drywell Calibrators ¹ | (-40 to 600) °C | 0.058 °C | Comparison to Platinum Resistance Thermometer, Multifunction Reference Thermometer |
| Temperature Baths ¹ | (-40 to 300) °C | 0.053 °C | Comparison to Platinum Resistance Thermometer, Multifunction Reference Thermometer |

Thermodynamic

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|---|--|
| RTD Probes ¹ (3 Wires & 4 Wire Sensors) | (-40 to 400) °C | 0.3 °C | Comparison to Drywell Calibrator, Platinum Resistance Thermometer, Multifunction Reference Thermometer |
| Infrared Thermometers | 50 °C 100 °C 200 °C 300 °C 390 °C | 0.62 °C 1.1 °C 1.5 °C 2.4 °C 3.3 °C | Comparison to Fluke 4181 Infrared Calibrator (flat plate) $\epsilon = 0.95, \lambda = (8 \text{ to } 14) \mu\text{m}$ |


Time and Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|----------------------------|---|--|---|
| Frequency – Source/Measure | 10 mHz to 119.99 Hz (120 to 1 199.9) Hz (1.2 to 11.999) kHz (12 to 119.99) kHz (120 to 1 199.9) kHz (1.2 to 2) MHz | 0.24 mHz 58 mHz 0.58 Hz 5.8 Hz 58 Hz 0.58 kHz | Comparison to Fluke 5522A Multiproduct Calibrator |

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. L = length in inches; DL = diagonal length in inches.
3. Unless otherwise specified in the far-right column, the calibration procedure/method utilized in the calibration of the listed parameter was written internally.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1255.



Jason Stine, Vice President